



Flexural Jig Assembly

Code: [39-6160](#)

Product Group: [Cement Testing Accessories](#)

Comes as standard with 39-7160/01 and 39-7160/06.

The Flexural Jig Assembly is designed for the Flexural Testing of 40 x 40 x 160 mm mortar prisms when fitted into ELE range of compression machines. The frame is supplied with a special ball seating assembly that is attached to the upper roller bearers.

Fitted with one upper and two lower roller bearers.