



Penetration/Swell Dial Gauge ASTM.

Code: 24-9184

Product Group: Expansion (swell) Test Equipment,
In-situ CBR, Conversion of In-situ to Laboratory CBR,
Penetration Measurement

1 in travel x 0.0005 in divisions. Complete with rack
extensions and chisel edge anvil.

Standards

BS 1377, BS 1924