

## Automatic Vicat Apparatus

Code: 38-2015/01

Product Group: Automatic Vicat

Complete with EN and ASTM Initial and Final needles, Consistency Plunger, Windows(R) software and RS232 cable, 1 x EN and ASTM Mould and Glass Plate. Suitable for continuous use in saturated humidity, at a controlled temperature of  $20^{\circ}\text{C} \pm 1^{\circ}\text{C}$ .

Complies to EN 13279-2 (GYPSUM), EN 480-2, ASTM C187, ASTM C191, DIN 1168, DIN 1196, NF P15-414, NF P15-431, AASHTO T131, EN 196-3

The results are printed on the incorporated printer and this eliminates the manual operations of installing and zeroing the paper graph on the drum. The use of the appliance is extremely simplified by the guiding menu that is available in english, french, spanish, german, italian language.

- Transfer each single control or function of the Vicat on the PC
- Verify in real time each phase of the test being made
- Automatically download the final results at the end of the test on all the connected Vicat
- Process and file at the same time all the tests without obliging the operator to move from his working place.

Additionally the firmware has many other functions detailed in the technical chart that will be transmitted to the user interested to know more about it.

---

## Standards

ASTM C187, ASTM C191

---


## Further Information

**\*\*NOT AVAILABLE FOR SALE IN THE USA\*\***

For 220 - 240 V AC, 50 - 60 Hz, 1 ph.

---


## Accessories



## ASTM Vicat Mould

Code: [38-2660](#)

---



## EN Vicat Mould

Code: [38-2300](#)

---



## Mould Tank

Code: [38-2015/15](#)

---



## Needle Cleaning Device

Code: [38-2015/12](#)

---



## Printer Paper Rolls

Code: [38-2015/14](#)

---

## Vicat Mould

Code: 38-2200



---

## Spares/Consumables



## ASTM Initial and Final Set Needle

Code: 38-2023



## Consistency Plunger

Code: 38-2015/16



## EN Initial and Final Set Needle

Code: 38-2021